

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060421\
 Data File : VY004974.D
 Acq On : 04 Jun 2021 19:00
 Operator : SY/MD
 Sample : M2613-13
 Misc : 10.84G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(134-138)

Quant Time: Jun 05 01:36:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	175709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	331490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	315718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	123654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	131137	70.597	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.200%
35) Dibromofluoromethane	7.728	113	106745	59.853	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.700%
50) Toluene-d8	10.185	98	415546	56.102	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.200%
62) 4-Bromofluorobenzene	12.483	95	134246	52.700	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.400%
Target Compounds						
16) Acetone	3.960	43	4998	11.177	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060421\
Data File : VY004974.D
Acq On : 04 Jun 2021 19:00
Operator : SY/MD
Sample : M2613-13
Misc : 10.84G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(134-138)

Quant Time: Jun 05 01:36:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

