

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060421\
 Data File : VY004975.D
 Acq On : 04 Jun 2021 19:23
 Operator : SY/MD
 Sample : M2612-01
 Misc : 8.18G/5ML/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(6-10)

Quant Time: Jun 05 01:36:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	182839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	346055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	336912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	149146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	128443	66.450	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.900%
35) Dibromofluoromethane	7.728	113	108953	58.520	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.040%
50) Toluene-d8	10.185	98	431625	55.820	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.640%
62) 4-Bromofluorobenzene	12.483	95	152383	57.302	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.600%
Target Compounds						
16) Acetone	3.954	43	6688	14.373	ug/l	Qvalue # 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060421\
Data File : VY004975.D
Acq On : 04 Jun 2021 19:23
Operator : SY/MD
Sample : M2612-01
Misc : 8.18G/5ML/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(6-10)

Quant Time: Jun 05 01:36:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

