

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060424\
 Data File : VY018460.D
 Acq On : 04 Jun 2024 13:39
 Operator : SY/MD
 Sample : P2701-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CON-PAD-1RE

Quant Time: Jun 05 01:24:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	83093	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.679	114	168048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	150446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	54330	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	84153	51.570	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.140%
35) Dibromofluoromethane	7.703	113	65220	30.606	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.220%
50) Toluene-d8	10.173	98	362374	42.745	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.500%
62) 4-Bromofluorobenzene	12.477	95	105466	39.957	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.920%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060424\
Data File : VY018460.D
Acq On : 04 Jun 2024 13:39
Operator : SY/MD
Sample : P2701-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CON-PAD-1RE

Quant Time: Jun 05 01:24:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY018460.D
Acq: 04 Jun 2024 13:39




Instrument :
MSVOA_Y
ClientSampleId :
CON-PAD-1RE


Tgt Ion:168 Resp: 83093
Ion Ratio Lower Upper
168 100
99 60.6 39.1 58.7#



#33
1,2-Dichloroethane-d4
Concen: 51.570 ug/l
RT: 8.130 min Scan# 1035
Delta R.T. -0.006 min
Lab File: VY018460.D
Acq: 04 Jun 2024 13:39

Tgt Ion: 65 Resp: 84153
Ion Ratio Lower Upper
65 100
67 53.5 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.679 min Scan# 1125

Delta R.T. -0.006 min

Lab File: VY018460.D

Acq: 04 Jun 2024 13:39

Instrument :

MSVOA_Y

ClientSampleId :

CON-PAD-1RE





Tgt Ion:114 Resp: 168048

Ion Ratio Lower Upper

114 100

63 20.5 0.0 35.0

88 14.7 0.0 27.2



#35

Dibromofluoromethane

Concen: 30.606 ug/l

RT: 7.703 min Scan# 965

Delta R.T. -0.000 min

Lab File: VY018460.D

Acq: 04 Jun 2024 13:39




Tgt Ion:113 Resp: 65220


Ion Ratio Lower Upper

113 100

111 103.6 82.2 123.4

192 16.6 15.9 23.9









#50
Toluene-d8
Concen: 42.745 ug/l
RT: 10.173 min Scan# 1370
Delta R.T. -0.000 min
Lab File: VY018460.D
Acq: 04 Jun 2024 13:39




Instrument :
MSVOA_Y
ClientSampleId :
CON-PAD-1RE


Tgt Ion: 98 Resp: 362374
Ion Ratio Lower Upper
98 100
100 65.5 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 39.957 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY018460.D
Acq: 04 Jun 2024 13:39

Tgt Ion: 95 Resp: 105466
Ion Ratio Lower Upper
95 100
174 74.9 0.0 175.6
176 73.8 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.483 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY018460.D

Acq: 04 Jun 2024 13:39

Instrument :

MSVOA_Y

ClientSampleId :

CON-PAD-1RE





Tgt Ion:117 Resp: 150446

Ion Ratio Lower Upper

117 100

82 57.1 42.4 63.6

119 31.5 25.9 38.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY018460.D

Acq: 04 Jun 2024 13:39

Tgt Ion:152 Resp: 54330

Ion Ratio Lower Upper

152 100

115 59.5 28.2 84.7

150 152.3 0.0 345.6

