

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060424\
 Data File : VY018466.D
 Acq On : 04 Jun 2024 16:06
 Operator : SY/MD
 Sample : P2687-04
 Misc : 3.64g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11977

Quant Time: Jun 05 01:27:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	65729	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.679	114	134513	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	122826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45941	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	76801	59.498	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.000%
35) Dibromofluoromethane	7.710	113	74832	43.871	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.740%
50) Toluene-d8	10.173	98	300213	44.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.480%
62) 4-Bromofluorobenzene	12.477	95	87277	41.309	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.620%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060424\
Data File : VY018466.D
Acq On : 04 Jun 2024 16:06
Operator : SY/MD
Sample : P2687-04
Misc : 3.64g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11977

Quant Time: Jun 05 01:27:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY018466.D
Acq: 04 Jun 2024 16:06




Instrument :
MSVOA_Y
ClientSampleId :
72-11977


Tgt Ion:168 Resp: 65729
Ion Ratio Lower Upper
168 100
99 59.8 39.1 58.7#



#33
1,2-Dichloroethane-d4
Concen: 59.498 ug/l
RT: 8.130 min Scan# 1035
Delta R.T. -0.006 min
Lab File: VY018466.D
Acq: 04 Jun 2024 16:06

Tgt Ion: 65 Resp: 76801
Ion Ratio Lower Upper
65 100
67 55.2 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.679 min Scan# 1125

Delta R.T. -0.006 min

Lab File: VY018466.D

Acq: 04 Jun 2024 16:06

Instrument :

MSVOA_Y

ClientSampleId :

72-11977





Tgt Ion:114 Resp: 134513

Ion Ratio Lower Upper

114 100

63 21.7 0.0 35.0

88 15.3 0.0 27.2



#35

Dibromofluoromethane

Concen: 43.871 ug/l

RT: 7.710 min Scan# 966

Delta R.T. 0.006 min

Lab File: VY018466.D

Acq: 04 Jun 2024 16:06




Tgt Ion:113 Resp: 74832


Ion Ratio Lower Upper

113 100

111 102.3 82.2 123.4

192 17.1 15.9 23.9









#50
Toluene-d8
Concen: 44.241 ug/l
RT: 10.173 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY018466.D
Acq: 04 Jun 2024 16:06




Instrument :
MSVOA_Y
ClientSampleId :
72-11977


Tgt Ion: 98 Resp: 300213
Ion Ratio Lower Upper
98 100
100 66.0 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 41.309 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY018466.D
Acq: 04 Jun 2024 16:06

Tgt Ion: 95 Resp: 87277
Ion Ratio Lower Upper
95 100
174 76.2 0.0 175.6
176 72.1 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.483 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY018466.D

Acq: 04 Jun 2024 16:06

Instrument :

MSVOA_Y

ClientSampleId :

72-11977





Tgt Ion:117 Resp: 122826

Ion Ratio Lower Upper

117 100

82 57.1 42.4 63.6

119 32.0 25.9 38.9



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY018466.D

Acq: 04 Jun 2024 16:06

Tgt Ion:152 Resp: 45941

Ion Ratio Lower Upper

152 100

115 58.3 28.2 84.7

150 155.8 0.0 345.6

