

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
 Data File : VY014006.D
 Acq On : 05 Jun 2023 14:46
 Operator : KP/MD
 Sample : 03016-04RE
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB29RE

Quant Time: Jun 06 03:54:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	4900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	9297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	10493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	3886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5988	92.315	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 184.640%#	
35) Dibromofluoromethane	7.716	113	3790	64.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.300%	
50) Toluene-d8	10.179	98	11358	50.397	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.800%	
62) 4-Bromofluorobenzene	12.483	95	3829	49.835	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.680%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	864	11.302	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
Data File : VY014006.D
Acq On : 05 Jun 2023 14:46
Operator : KP/MD
Sample : 03016-04RE
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB29RE

Quant Time: Jun 06 03:54:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

