

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
 Data File : VY014011.D
 Acq On : 05 Jun 2023 16:43
 Operator : KP/MD
 Sample : 03018-05
 Misc : 3.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 348

Quant Time: Jun 06 03:57:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	156916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	272077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	275211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	120010	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	116205	55.943	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.880%
35) Dibromofluoromethane	7.716	113	93775	54.234	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.460%
50) Toluene-d8	10.179	98	340259	51.590	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.180%
62) 4-Bromofluorobenzene	12.477	95	110226	49.022	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
Data File : VY014011.D
Acq On : 05 Jun 2023 16:43
Operator : KP/MD
Sample : 03018-05
Misc : 3.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
348

Quant Time: Jun 06 03:57:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

