

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
 Data File : VY014017.D
 Acq On : 05 Jun 2023 19:03
 Operator : KP/MD
 Sample : 02976-01
 Misc : 6.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Jun 06 03:59:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	154808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	269194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	278263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	131990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	110640	53.989	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.980%
35) Dibromofluoromethane	7.716	113	92894	54.300	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.600%
50) Toluene-d8	10.179	98	336195	51.520	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.040%
62) 4-Bromofluorobenzene	12.483	95	117985	53.034	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060523\
Data File : VY014017.D
Acq On : 05 Jun 2023 19:03
Operator : KP/MD
Sample : 02976-01
Misc : 6.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Jun 06 03:59:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

