

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
 Data File : VY018485.D
 Acq On : 05 Jun 2024 15:22
 Operator : SY/MD
 Sample : P2680-03RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-241RE

Quant Time: Jun 05 15:47:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	62364	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	131034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	119867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	42753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	71940	58.739	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.480%
35) Dibromofluoromethane	7.710	113	67754	40.776	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.560%
50) Toluene-d8	10.173	98	170601	25.808	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	51.620%#
62) 4-Bromofluorobenzene	12.477	95	61958	30.104	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	60.200%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.698	84	8401	2.665	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
Data File : VY018485.D
Acq On : 05 Jun 2024 15:22
Operator : SY/MD
Sample : P2680-03RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-241RE

Quant Time: Jun 05 15:47:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

