

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
 Data File : VY018488.D
 Acq On : 05 Jun 2024 16:35
 Operator : SY/MD
 Sample : P2715-03
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WCS-TPC-VOC

Quant Time: Jun 06 05:57:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	58738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	121047	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	110341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	39955	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	66852	57.955	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.900%
35) Dibromofluoromethane	7.710	113	63831	41.585	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.160%
50) Toluene-d8	10.173	98	260466	42.654	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.300%
62) 4-Bromofluorobenzene	12.477	95	71348	37.526	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
Data File : VY018488.D
Acq On : 05 Jun 2024 16:35
Operator : SY/MD
Sample : P2715-03
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WCS-TPC-VOC

Quant Time: Jun 06 05:57:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

