

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
 Data File : VY018493.D
 Acq On : 05 Jun 2024 18:39
 Operator : SY/MD
 Sample : P2683-04
 Misc : 7.23g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-5-VOC

Quant Time: Jun 06 05:59:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	55248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	109296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	99421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	35843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	60778	56.017	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.040%
35) Dibromofluoromethane	7.710	113	57365	41.390	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.780%
50) Toluene-d8	10.173	98	228558	41.453	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.900%
62) 4-Bromofluorobenzene	12.477	95	59724	34.790	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	69.580%
Target Compounds						
20) Methylene Chloride	4.698	84	14945	8.288	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
Data File : VY018493.D
Acq On : 05 Jun 2024 18:39
Operator : SY/MD
Sample : P2683-04
Misc : 7.23g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-5-VOC

Quant Time: Jun 06 05:59:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

