

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
 Data File : VY018496.D
 Acq On : 05 Jun 2024 19:52
 Operator : SY/MD
 Sample : P2668-07
 Misc : 5.35g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC-TP-2

Quant Time: Jun 06 06:00:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	6403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	11693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	8124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	5481	43.588	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.180%
35) Dibromofluoromethane	7.704	113	4539	30.612	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.220%
50) Toluene-d8	10.173	98	15859	26.885	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	53.780%#
62) 4-Bromofluorobenzene	12.477	95	3317	18.060	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	36.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
Data File : VY018496.D
Acq On : 05 Jun 2024 19:52
Operator : SY/MD
Sample : P2668-07
Misc : 5.35g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC-TP-2

Quant Time: Jun 06 06:00:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

