



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060524\
Data File : VY018479.D
Acq On : 05 Jun 2024 11:38
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Title : SW846 8260
Last Update : Tue Jun 04 07:11:56 2024



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	15805	PASS
75	95	30	60	56.9	38653	PASS
95	95	100	100	100.0	67949	PASS
96	95	5	9	6.8	4608	PASS
173	174	0.00	2	0.9	463	PASS
174	95	50	100	74.5	50653	PASS
175	174	5	9	6.8	3423	PASS
176	174	95	101	96.5	48885	PASS
177	176	5	9	6.6	3204	PASS