

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060624\
 Data File : VY018510.D
 Acq On : 06 Jun 2024 17:15
 Operator : SY/MD
 Sample : P2707-04
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Jun 07 02:28:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	87	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	31	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.379	152	25	50.000	ug/l	#-0.04

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	196	114.718	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 229.440%#	
35) Dibromofluoromethane	7.697	113	115	35.015	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 70.040%	
50) Toluene-d8	10.179	98	343	26.252	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 52.500%#	
62) 4-Bromofluorobenzene	12.642	95	31	7.620	ug/l	0.16
Spiked Amount	50.000	Range	30 - 143	Recovery	= 15.240%#	

Target Compounds			Qvalue			
2) Dichlorodifluoromethane	1.887	85	26	18.817	ug/l #	43
3) Chloromethane	2.113	50	30	11.642	ug/l #	43
4) Vinyl Chloride	2.265	62	25	8.111	ug/l #	42
5) Bromomethane	2.625	94	43	20.911	ug/l #	3
6) Chloroethane	2.771	64	23	11.819	ug/l #	11
7) Trichlorofluoromethane	2.887	101	33	10.348	ug/l #	20
8) Diethyl Ether	3.637	74	88	91.713	ug/l	64
10) Methyl Iodide	4.076	142	22	12.062	ug/l #	1
11) Tert butyl alcohol	5.021	59	18	39.224	ug/l #	1
12) 1,1-Dichloroethene	3.960	96	52	28.491	ug/l #	1
13) Acrolein	3.643	56	41	234.441	ug/l #	1
14) Allyl chloride	4.448	41	46	17.209	ug/l #	19
15) Acrylonitrile	5.143	53	28	63.810	ug/l #	1
16) Acetone	3.912	43	138	363.913	ug/l #	36
17) Carbon Disulfide	4.180	76	26	4.856	ug/l #	76
18) Methyl Acetate	4.460	43	46	48.509	ug/l #	46
19) Methyl tert-butyl Ether	5.015	73	21	4.774	ug/l #	54
20) Methylene Chloride	4.674	84	27	9.958	ug/l #	45
21) trans-1,2-Dichloroethene	4.978	96	25	12.836	ug/l #	1
22) Diisopropyl ether	6.125	45	67	12.308	ug/l #	37
23) Vinyl Acetate	6.027	43	107	30.203	ug/l #	67
24) 1,1-Dichloroethane	5.984	63	48	14.137	ug/l #	80
25) 2-Butanone	6.960	43	85	146.316	ug/l #	41
26) 2,2-Dichloropropane	7.021	77	21	7.037	ug/l #	51
27) cis-1,2-Dichloroethene	6.960	96	26	11.476	ug/l #	35
28) Bromochloromethane	7.338	49	27	17.280	ug/l #	6
29) Tetrahydrofuran	7.350	42	21	55.662	ug/l #	26
30) Chloroform	7.441	83	21	6.167	ug/l #	17
31) Cyclohexane	7.783	56	133	41.751	ug/l #	11
32) 1,1,1-Trichloroethane	7.691	97	42	14.081	ug/l #	25
36) 1,1-Dichloropropene	7.947	75	25	4.992	ug/l #	40
37) Ethyl Acetate	7.015	43	26	10.393	ug/l #	1
38) Carbon Tetrachloride	8.149	117	20	3.985	ug/l #	13
39) Methylcyclohexane	9.112	83	33	4.875	ug/l #	22
40) Benzene	8.161	78	43	2.866	ug/l #	51
41) Methacrylonitrile	7.234	41	40	32.871	ug/l #	1

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060624\
 Data File : VY018510.D
 Acq On : 06 Jun 2024 17:15
 Operator : SY/MD
 Sample : P2707-04
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :


Quant Time: Jun 07 02:28:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

42) 1,2-Dichloroethane	8.283	62	58	15.900	ug/l #	71
43) Isopropyl Acetate	8.283	43	76	16.247	ug/l #	47
45) 1,2-Dichloropropane	9.209	63	54	15.737	ug/l #	43
46) Dibromomethane	9.228	93	19	9.657	ug/l #	3
47) Bromodichloromethane	9.471	83	19	4.023	ug/l #	26
48) Methyl methacrylate	9.307	41	51	23.697	ug/l #	9
49) 1,4-Dioxane	9.508	88	21	812.824	ug/l #	1
51) 4-Methyl-2-Pentanone	10.112	43	66	27.172	ug/l	97
52) Toluene	10.191	92	50	5.418	ug/l #	1
54) cis-1,3-Dichloropropene	9.801	75	23	4.307	ug/l #	61
56) Ethyl methacrylate	10.496	69	20	5.553	ug/l #	33
58) 2-Chloroethyl Vinyl ether	9.971	63	34	17.578	ug/l #	45
59) 2-Hexanone	10.758	43	133	79.236	ug/l	96
60) Dibromochloromethane	10.709	129	34	10.873	ug/l	78
65) Chlorobenzene	11.331	112	18	13.540	ug/l #	42
67) Ethyl Benzene	11.776	91	26	10.745	ug/l #	44
68) m/p-Xylenes	11.605	106	45	49.389	ug/l #	1
69) o-Xylene	12.062	106	44	50.166	ug/l #	1
71) Bromoform	12.294	173	25	110.396	ug/l #	27
73) Isopropylbenzene	12.526	105	42	16.280	ug/l #	48
74) N-amyl acetate	12.136	43	63	100.325	ug/l #	37
75) 1,1,2,2-Tetrachloroethane	12.587	83	27	59.454	ug/l #	26
76) 1,2,3-Trichloropropane	12.611	75	20	64.516	ug/l #	100
78) n-propylbenzene	12.611	91	19	6.136	ug/l #	53
79) 2-Chlorotoluene	12.611	91	19	11.287	ug/l #	39
80) 1,3,5-Trimethylbenzene	12.526	105	42	20.903	ug/l #	27
81) trans-1,4-Dichloro-2-b...	12.459	75	105	711.600	ug/l #	16
82) 4-Chlorotoluene	12.611	91	19	11.217	ug/l #	39
84) 1,2,4-Trimethylbenzene	13.276	105	30	15.251	ug/l #	30
85) sec-Butylbenzene	13.276	105	30	10.964	ug/l #	56
90) Hexachloroethane	14.062	117	24	57.980	ug/l #	12
92) 1,2-Dibromo-3-Chloropr...	14.355	75	61	882.691	ug/l #	1
93) 1,2,4-Trichlorobenzene	14.946	180	21	41.298	ug/l #	12
95) Naphthalene	15.397	128	58	56.793	ug/l #	1
96) 1,2,3-Trichlorobenzene	15.184	180	63	146.430	ug/l #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 07 02:28:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 87
Ion Ratio Lower Upper
168 100
99 0.0 39.1 58.7#



#2
Dichlorodifluoromethane
Concen: 18.817 ug/l
RT: 1.887 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 85 Resp: 26
Ion Ratio Lower Upper
85 100
87 0.0 16.0 48.0#









#3
Chloromethane
Concen: 11.642 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.012 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 50 Resp: 30
Ion Ratio Lower Upper
50 100
52 0.0 25.6 38.4#



#4
Vinyl Chloride
Concen: 8.111 ug/l
RT: 2.265 min Scan# 73
Delta R.T. 0.024 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 62 Resp: 25
Ion Ratio Lower Upper
62 100
64 0.0 26.1 39.1#









#5
Bromomethane
Concen: 20.911 ug/l
RT: 2.625 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 94 Resp: 43
Ion Ratio Lower Upper
94 100
96 0.0 74.1 111.1#



#6
Chloroethane
Concen: 11.819 ug/l
RT: 2.771 min Scan# 156
Delta R.T. -0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 64 Resp: 23
Ion Ratio Lower Upper
64 100
66 82.5 26.1 39.1#









#7
Trichlorofluoromethane
Concen: 10.348 ug/l
RT: 2.887 min Scan# 11
Delta R.T. -0.220 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 33
Ion Ratio Lower Upper
101 100
103 0.0 49.3 73.9#



#8
Diethyl Ether
Concen: 91.713 ug/l
RT: 3.637 min Scan# 298
Delta R.T. 0.122 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 74 Resp: 88
Ion Ratio Lower Upper
74 100
45 50.0 41.3 124.1









#10
Methyl Iodide
Concen: 12.062 ug/l
RT: 4.076 min Scan# 31
Delta R.T. 0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:142 Resp: 22
Ion Ratio Lower Upper
142 100
127 209.1 32.8 49.2#
141 0.0 11.4 17.2#



#11
Tert butyl alcohol
Concen: 39.224 ug/l
RT: 5.021 min Scan# 525
Delta R.T. 0.085 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 59 Resp: 18
Ion Ratio Lower Upper
59 100
57 105.6 3.0 4.6#





#12

1,1-Dichloroethene

Concen: 28.491 ug/l

RT: 3.960 min Scan# 311

Delta R.T. 0.098 min

Lab File: VY018510.D





Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion: 96	Resp:	52
Ion	Ratio	Lower Upper
96	100	
61	0.0	114.6 171.8#
98	0.0	50.8 76.2#



#13

Acrolein

Concen: 234.441 ug/l




RT: 3.643 min Scan# 299


Delta R.T. -0.073 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15


Tgt Ion: 56	Resp:	41
Ion	Ratio	Lower Upper
56	100	
55	265.9	55.7 83.5#








#14
 Allyl chloride
 Concen: 17.209 ug/l
 RT: 4.448 min Scan# 41
 Delta R.T. -0.018 min
 Lab File: VY018510.D
 Acq: 06 Jun 2024 17:15

Instrument :
 MSVOA_Y
 ClientSampleId :







Tgt Ion: 41 Resp: 46
 Ion Ratio Lower Upper
 41 100
 39 0.0 58.3 87.5#
 76 0.0 39.0 58.6#



#15
 Acrylonitrile
 Concen: 63.810 ug/l
 RT: 5.143 min Scan# 545
 Delta R.T. -0.000 min
 Lab File: VY018510.D
 Acq: 06 Jun 2024 17:15

Tgt Ion: 53 Resp: 28
 Ion Ratio Lower Upper
 53 100
 52 382.1 64.7 97.1#
 51 0.0 28.4 42.6#









#16
Acetone
Concen: 363.913 ug/l
RT: 3.912 min Scan# 347
Delta R.T. -0.024 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 138
Ion Ratio Lower Upper
43 100
58 0.0 31.7 47.5#



#17
Carbon Disulfide
Concen: 4.856 ug/l
RT: 4.180 min Scan# 387
Delta R.T. -0.000 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Tgt Ion: 76 Resp: 26
Ion Ratio Lower Upper
76 100
78 0.0 6.8 10.2#








#18
Methyl Acetate
Concen: 48.509 ug/l
RT: 4.460 min Scan# 41
Delta R.T. -0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 46
Ion Ratio Lower Upper
43 100
74 0.0 23.3 34.9#



#19
Methyl tert-butyl Ether
Concen: 4.774 ug/l
RT: 5.015 min Scan# 524
Delta R.T. -0.195 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Tgt Ion: 73 Resp: 21
Ion Ratio Lower Upper
73 100
57 0.0 17.3 25.9#








#20
Methylene Chloride
Concen: 9.958 ug/l
RT: 4.674 min Scan# 41
Delta R.T. -0.024 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:	84	Resp:	27
Ion	Ratio	Lower	Upper
84	100		
49	145.3	84.4	126.6#
51	0.0	26.0	39.0#
86	0.0	49.5	74.3#



#21
trans-1,2-Dichloroethene
Concen: 12.836 ug/l
RT: 4.978 min Scan# 518
Delta R.T. -0.220 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Tgt Ion:	96	Resp:	25
Ion	Ratio	Lower	Upper
96	100		
61	0.0	97.5	146.3#
98	0.0	50.5	75.7#








#22
Diisopropyl ether
Concen: 12.308 ug/l
RT: 6.125 min Scan# 703
Delta R.T. 0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 45	Resp: 67
Ion Ratio	Lower Upper
45	100
43	0.0 42.2 63.2#
87	0.0 25.0 37.6#
59	0.0 10.5 15.7#



#23
Vinyl Acetate
Concen: 30.203 ug/l
RT: 6.027 min Scan# 690
Delta R.T. -0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 43	Resp: 107
Ion Ratio	Lower Upper
43	100
86	0.0 10.6 16.0#









#24
1,1-Dichloroethane
Concen: 14.137 ug/l
RT: 5.984 min Scan# 61
Delta R.T. -0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 63 Resp: 48
Ion Ratio Lower Upper
63 100
98 0.0 4.0 12.2#
100 0.0 2.6 8.0#



#25
2-Butanone
Concen: 146.316 ug/l
RT: 6.960 min Scan# 843
Delta R.T. -0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 43 Resp: 85
Ion Ratio Lower Upper
43 100
72 0.0 27.3 40.9#





#26

2,2-Dichloropropane

Concen: 7.037 ug/l

RT: 7.021 min Scan# 8

Delta R.T. 0.049 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 77 Resp: 21

Ion Ratio Lower Upper

77 100

97 0.0 12.0 36.1#



#27

cis-1,2-Dichloroethene

Concen: 11.476 ug/l

RT: 6.960 min Scan# 843

Delta R.T. -0.012 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15




Tgt Ion: 96 Resp: 26


Ion Ratio Lower Upper

96 100

61 192.3 0.0 256.2

98 0.0 0.0 130.0









#28
Bromochloromethane
Concen: 17.280 ug/l
RT: 7.338 min Scan# 905
Delta R.T. 0.012 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 49 Resp: 27
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.6
130 0.0 72.7 109.1#



#29
Tetrahydrofuran
Concen: 55.662 ug/l
RT: 7.350 min Scan# 907
Delta R.T. 0.012 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 42 Resp: 21
Ion Ratio Lower Upper
42 100
72 0.0 43.0 64.4#
71 0.0 40.8 61.2#









#30
Chloroform
Concen: 6.167 ug/l
RT: 7.441 min Scan# 91
Delta R.T. -0.055 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 21
Ion Ratio Lower Upper
83 100
85 0.0 52.2 78.4#



#31
Cyclohexane
Concen: 41.751 ug/l
RT: 7.783 min Scan# 978
Delta R.T. 0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 56 Resp: 133
Ion Ratio Lower Upper
56 100
69 0.0 28.7 43.1#
84 0.0 77.7 116.5#









#32
1,1,1-Trichloroethane
Concen: 14.081 ug/l
RT: 7.691 min Scan# 90
Delta R.T. -0.000 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 97 Resp: 42
Ion Ratio Lower Upper
97 100
99 0.0 50.3 75.5#
61 0.0 32.6 48.8#



#33
1,2-Dichloroethane-d4
Concen: 114.718 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.000 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 65 Resp: 196
Ion Ratio Lower Upper
65 100
67 51.0 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:114 Resp: 259

Ion Ratio Lower Upper

114 100

63 31.5 0.0 35.0

88 0.0 0.0 27.2



#35

Dibromofluoromethane

Concen: 35.015 ug/l

RT: 7.697 min Scan# 964

Delta R.T. -0.006 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15




Tgt Ion:113 Resp: 115


Ion Ratio Lower Upper

113 100

111 107.8 82.2 123.4


192 0.0 15.9 23.9#








#36
1,1-Dichloropropene
Concen: 4.992 ug/l
RT: 7.947 min Scan# 1005
Delta R.T. 0.037 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 75 Resp: 25
Ion Ratio Lower Upper
75 100
110 0.0 19.0 57.0#
77 0.0 25.3 37.9#



#37
Ethyl Acetate
Concen: 10.393 ug/l
RT: 7.015 min Scan# 852
Delta R.T. -0.043 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Tgt Ion: 43 Resp: 26
Ion Ratio Lower Upper
43 100
61 146.2 12.6 19.0#
70 338.5 7.8 11.8#








#38
Carbon Tetrachloride
Concen: 3.985 ug/l
RT: 8.149 min Scan# 1196
Delta R.T. 0.256 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:117 Resp: 20
Ion Ratio Lower Upper
117 100
119 0.0 75.8 113.6#
121 0.0 24.0 36.0#



#39
Methylcyclohexane
Concen: 4.875 ug/l
RT: 9.112 min Scan# 1196
Delta R.T. -0.061 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 83 Resp: 33
Ion Ratio Lower Upper
83 100
55 0.0 51.5 77.3#
98 0.0 40.5 60.7#









#40
Benzene
Concen: 2.866 ug/l
RT: 8.161 min Scan# 1040
Delta R.T. 0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 78 Resp: 43
Ion Ratio Lower Upper
78 100
77 0.0 19.5 29.3#



#41
Methacrylonitrile
Concen: 32.871 ug/l
RT: 7.234 min Scan# 888
Delta R.T. -0.061 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 41 Resp: 40
Ion Ratio Lower Upper
41 100
39 0.0 48.0 72.0#
67 0.0 75.1 112.7#
52 140.0 29.8 44.6#





#42

1,2-Dichloroethane

Concen: 15.900 ug/l

RT: 8.283 min Scan# 1060

Delta R.T. 0.055 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 62 Resp: 58

Ion Ratio Lower Upper

62 100

98 0.0 0.0 22.0



#43

Isopropyl Acetate

Concen: 16.247 ug/l

RT: 8.283 min Scan# 1060

Delta R.T. 0.018 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15




Tgt Ion: 43 Resp: 76


Ion Ratio Lower Upper

43 100

61 0.0 27.9 41.9#

87 0.0 13.1 19.7#





#45

1,2-Dichloropropane

Concen: 15.737 ug/l

RT: 9.209 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 63 Resp: 54

Ion Ratio Lower Upper

63 100

65 0.0 25.4 38.2#



#46

Dibromomethane

Concen: 9.657 ug/l

RT: 9.228 min Scan# 1215

Delta R.T. -0.067 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15




Tgt Ion: 93 Resp: 19


Ion Ratio Lower Upper

93 100

95 0.0 66.2 99.4#


174 0.0 80.9 121.3#








#47
Bromodichloromethane
Concen: 4.023 ug/l
RT: 9.471 min Scan# 11
Delta R.T. -0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 83 Resp: 19
Ion Ratio Lower Upper
83 100
85 0.0 50.1 75.1#
127 0.0 7.4 11.0#



#48
Methyl methacrylate
Concen: 23.697 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. 0.018 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 41 Resp: 51
Ion Ratio Lower Upper
41 100
69 0.0 84.1 126.1#
39 103.9 43.6 65.4#









#49
1,4-Dioxane
Concen: 812.824 ug/l
RT: 9.508 min Scan# 11
Delta R.T. 0.207 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 88 Resp: 21
Ion Ratio Lower Upper
88 100
43 0.0 0.0 0.0
58 304.8 49.3 73.9#



#50
Toluene-d8
Concen: 26.252 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. 0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 98 Resp: 343
Ion Ratio Lower Upper
98 100
100 52.2 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 27.172 ug/l
RT: 10.112 min Scan# 11
Delta R.T. 0.049 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 66
Ion Ratio Lower Upper
43 100
58 42.4 35.7 53.5



#52
Toluene
Concen: 5.418 ug/l
RT: 10.191 min Scan# 1373
Delta R.T. -0.049 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Tgt Ion: 92 Resp: 50
Ion Ratio Lower Upper
92 100
91 0.0 139.0 208.6#








#54
cis-1,3-Dichloropropene
Concen: 4.307 ug/l
RT: 9.801 min Scan# 11
Delta R.T. -0.122 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 75 Resp: 23
Ion Ratio Lower Upper
75 100
77 0.0 25.4 38.0#
39 22.6 29.8 44.6#



#56
Ethyl methacrylate
Concen: 5.553 ug/l
RT: 10.496 min Scan# 1423
Delta R.T. -0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 69 Resp: 20
Ion Ratio Lower Upper
69 100
41 0.0 42.3 63.5#
39 0.0 22.9 34.3#









#58
2-Chloroethyl Vinyl ether
Concen: 17.578 ug/l
RT: 9.971 min Scan# 11
Delta R.T. 0.195 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 63 Resp: 34
Ion Ratio Lower Upper
63 100
106 0.0 23.7 35.5#



#59
2-Hexanone
Concen: 79.236 ug/l
RT: 10.758 min Scan# 1466
Delta R.T. -0.067 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 43 Resp: 133
Ion Ratio Lower Upper
43 100
58 59.4 31.4 94.0









#60
Dibromochloromethane
Concen: 10.873 ug/l
RT: 10.709 min Scan# 1458
Delta R.T. -0.268 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:129 Resp: 34
Ion Ratio Lower Upper
129 100
127 58.8 39.0 117.0



#62
4-Bromofluorobenzene
Concen: 7.620 ug/l
RT: 12.642 min Scan# 1775
Delta R.T. 0.165 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 95 Resp: 31
Ion Ratio Lower Upper
95 100
174 0.0 0.0 175.6
176 0.0 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.496 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY018510.D





Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion:117	Resp:	31
Ion	Ratio	Lower Upper
117	100	
82	181.4	42.4 63.6#
119	0.0	25.9 38.9#



#65

Chlorobenzene

Concen: 13.540 ug/l




RT: 11.331 min Scan# 1560


Delta R.T. -0.183 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Tgt Ion:112	Resp:	18
Ion	Ratio	Lower Upper
112	100	
114	0.0	26.4 39.6#









#67
Ethyl Benzene
Concen: 10.745 ug/l
RT: 11.776 min Scan# 11
Delta R.T. 0.189 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 26
Ion Ratio Lower Upper
91 100
106 0.0 24.5 36.7#



#68
m/p-Xylenes
Concen: 49.389 ug/l
RT: 11.605 min Scan# 1605
Delta R.T. -0.091 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion:106 Resp: 45
Ion Ratio Lower Upper
106 100
91 0.0 156.0 234.0#









#69
o-Xylene
Concen: 50.166 ug/l
RT: 12.062 min Scan# 11
Delta R.T. 0.037 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:106 Resp: 44
Ion Ratio Lower Upper
106 100
91 0.0 103.6 310.8#



#71
Bromoform
Concen: 110.396 ug/l
RT: 12.294 min Scan# 1718
Delta R.T. 0.091 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion:173 Resp: 25
Ion Ratio Lower Upper
173 100
175 0.0 25.1 75.3#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.379 min Scan# 11

Delta R.T. -0.043 min


Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion:152 Resp: 25

Ion Ratio Lower Upper

152 100

115 0.0 28.2 84.7#

150 0.0 0.0 345.6



#73

Isopropylbenzene

Concen: 16.280 ug/l

RT: 12.526 min Scan# 1756

Delta R.T. 0.201 min

Lab File: VY018510.D




Acq: 06 Jun 2024 17:15


Tgt Ion:105 Resp: 42

Ion Ratio Lower Upper

105 100


120 0.0 13.4 40.1#








#74
N-amy1 acetate
Concen: 100.325 ug/l
RT: 12.136 min Scan# 1692
Delta R.T. -0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43	Resp: 63
Ion Ratio	Lower Upper
43	100
70	0.0 41.4 62.2#
55	0.0 23.3 34.9#
61	0.0 22.6 33.8#



#75
1,1,2,2-Tetrachloroethane
Concen: 59.454 ug/l
RT: 12.587 min Scan# 1766
Delta R.T. 0.006 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 83	Resp: 27
Ion Ratio	Lower Upper
83	100
131	0.0 5.0 14.9#
85	0.0 32.0 96.0#





#76

1,2,3-Trichloropropane

Concen: 64.516 ug/l

RT: 12.611 min Scan# 1770

Delta R.T. -0.018 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 75 Resp: 20

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0



#78

n-propylbenzene

Concen: 6.136 ug/l

RT: 12.611 min Scan# 1770

Delta R.T. -0.055 min

Lab File: VY018510.D




Acq: 06 Jun 2024 17:15


Tgt Ion: 91 Resp: 19

Ion Ratio Lower Upper

91 100

120 0.0 11.5 34.4#









#79
2-Chlorotoluene
Concen: 11.287 ug/l
RT: 12.611 min Scan# 1756
Delta R.T. -0.140 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 19
Ion Ratio Lower Upper
91 100
126 0.0 17.8 53.5#



#80
1,3,5-Trimethylbenzene
Concen: 20.903 ug/l
RT: 12.526 min Scan# 1756
Delta R.T. -0.280 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 105 Resp: 42
Ion Ratio Lower Upper
105 100
120 0.0 25.1 75.2#









#81
trans-1,4-Dichloro-2-butene
Concen: 711.600 ug/l
RT: 12.459 min Scan# 11
Delta R.T. 0.085 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 105
Ion Ratio Lower Upper
75 100
53 0.0 66.8 100.2#
89 0.0 35.8 53.6#



#82
4-Chlorotoluene
Concen: 11.217 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. -0.238 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Tgt Ion: 91 Resp: 19
Ion Ratio Lower Upper
91 100
126 0.0 17.6 52.9#





#84

1,2,4-Trimethylbenzene

Concen: 15.251 ug/l

RT: 13.276 min Scan# 1879

Delta R.T. 0.158 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 30

Ion Ratio Lower Upper

105 100

120 0.0 23.2 69.6#



#85

sec-Butylbenzene

Concen: 10.964 ug/l

RT: 13.276 min Scan# 1879

Delta R.T. 0.024 min

Lab File: VY018510.D




Acq: 06 Jun 2024 17:15


Tgt Ion:105 Resp: 30

Ion Ratio Lower Upper

105 100

134 0.0 10.1 30.1#





#90

Hexachloroethane

Concen: 57.980 ug/l

RT: 14.062 min Scan# 2008

Delta R.T. 0.104 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:117 Resp: 24

Ion Ratio Lower Upper

117 100

201 0.0 37.4 112.2#



#92

1,2-Dibromo-3-Chloropropane

Concen: 882.691 ug/l

RT: 14.355 min Scan# 2056

Delta R.T. 0.006 min

Lab File: VY018510.D

Acq: 06 Jun 2024 17:15




Tgt Ion: 75 Resp: 61


Ion Ratio Lower Upper

75 100

155 0.0 48.6 145.8#


157 0.0 62.7 188.2#








#93
1,2,4-Trichlorobenzene
Concen: 41.298 ug/l
RT: 14.946 min Scan# 2162
Delta R.T. -0.055 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion:180 Resp: 21
Ion Ratio Lower Upper
180 100
182 0.0 47.9 143.8#
145 0.0 14.2 42.6#




#95
Naphthalene
Concen: 56.793 ug/l
RT: 15.397 min Scan# 2227
Delta R.T. 0.165 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15




Tgt Ion:128 Resp: 58
Ion Ratio Lower Upper
128 100
127 0.0 10.6 16.0#
129 86.2 8.6 13.0#





#96
1,2,3-Trichlorobenzene
Concen: 146.430 ug/l
RT: 15.184 min Scan# 21
Delta R.T. -0.238 min
Lab File: VY018510.D
Acq: 06 Jun 2024 17:15

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.6	142.9#
145	0.0	15.4	46.1#

