

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060624\
 Data File : VY018516.D
 Acq On : 06 Jun 2024 19:43
 Operator : SY/MD
 Sample : P2740-07
 Misc : 5.26g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WCS-TPB-VOC

Quant Time: Jun 07 02:32:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	50209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	108977	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	102186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	37574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	65289	66.214	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.420%
35) Dibromofluoromethane	7.704	113	58152	42.081	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.160%
50) Toluene-d8	10.173	98	240369	43.723	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.440%
62) 4-Bromofluorobenzene	12.477	95	68028	39.743	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.480%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.698	84	12977	7.785	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060624\
Data File : VY018516.D
Acq On : 06 Jun 2024 19:43
Operator : SY/MD
Sample : P2740-07
Misc : 5.26g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WCS-TPB-VOC

Quant Time: Jun 07 02:32:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

