

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
 Data File : VY004990.D
 Acq On : 07 Jun 2021 14:13
 Operator : SY/MD
 Sample : M2613-12
 Misc : 8.44G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(130-134)

Quant Time: Jun 08 02:08:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	18003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	25092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	18772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	6740	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	6655	34.967	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	69.940%
35) Dibromofluoromethane	7.722	113	4434	32.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	65.680%
50) Toluene-d8	10.185	98	29812	53.173	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.340%
62) 4-Bromofluorobenzene	12.483	95	6310	32.724	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
Data File : VY004990.D
Acq On : 07 Jun 2021 14:13
Operator : SY/MD
Sample : M2613-12
Misc : 8.44G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(130-134)

Quant Time: Jun 08 02:08:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

