

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
 Data File : VY004995.D
 Acq On : 07 Jun 2021 16:09
 Operator : SY/MD
 Sample : M2612-12
 Misc : 10.41G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(50-54)

Quant Time: Jun 08 02:09:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	185032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	352721	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	347402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	157493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	124967	63.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.780%	
35) Dibromofluoromethane	7.728	113	108910	57.391	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.780%	
50) Toluene-d8	10.185	98	443349	56.253	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.500%	
62) 4-Bromofluorobenzene	12.483	95	150366	55.475	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.940%	
Target Compounds						Qvalue
16) Acetone	3.954	43	5087	10.803	ug/l	99
17) Carbon Disulfide	4.204	76	13272	2.270	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
Data File : VY004995.D
Acq On : 07 Jun 2021 16:09
Operator : SY/MD
Sample : M2612-12
Misc : 10.41G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(50-54)

Quant Time: Jun 08 02:09:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

