

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
 Data File : VY004996.D
 Acq On : 07 Jun 2021 16:32
 Operator : SY/MD
 Sample : M2612-13
 Misc : 6.19G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(54-58)

Quant Time: Jun 08 02:09:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	177839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	334172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	330851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	152077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	122762	65.296	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	130.600%	
35) Dibromofluoromethane	7.728	113	105281	58.558	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	117.120%	
50) Toluene-d8	10.185	98	420898	56.369	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	112.740%	
62) 4-Bromofluorobenzene	12.483	95	143857	56.020	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	112.040%	
Target Compounds						Qvalue
16) Acetone	3.960	43	12229	27.021	ug/l	99
17) Carbon Disulfide	4.198	76	31012	5.518	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060721\
Data File : VY004996.D
Acq On : 07 Jun 2021 16:32
Operator : SY/MD
Sample : M2612-13
Misc : 6.19G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(54-58)

Quant Time: Jun 08 02:09:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

