

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060722\
 Data File : VY009159.D
 Acq On : 07 Jun 2022 12:01
 Operator : KP/MD
 Sample : VY0607SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0607SBL01

Quant Time: Jun 07 17:22:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	106565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	195780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	181513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	68519	62.475	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.940%
35) Dibromofluoromethane	7.716	113	62631	51.445	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.880%
50) Toluene-d8	10.179	98	236184	52.587	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.180%
62) 4-Bromofluorobenzene	12.483	95	68463	37.788	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060722\
Data File : VY009159.D
Acq On : 07 Jun 2022 12:01
Operator : KP/MD
Sample : VY0607SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0607SBL01

Quant Time: Jun 07 17:22:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

