

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
 Data File : VY014059.D
 Acq On : 07 Jun 2023 14:55
 Operator : KP/MD
 Sample : 03069-07
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB36

Quant Time: Jun 08 02:33:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	140943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	236577	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	232872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	105296	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	87139	46.704	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.400%
35) Dibromofluoromethane	7.722	113	78353	52.115	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.220%
50) Toluene-d8	10.185	98	297326	51.845	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.700%
62) 4-Bromofluorobenzene	12.483	95	91661	46.882	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
Data File : VY014059.D
Acq On : 07 Jun 2023 14:55
Operator : KP/MD
Sample : 03069-07
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB36

Quant Time: Jun 08 02:33:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

