

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
 Data File : VY014060.D
 Acq On : 07 Jun 2023 15:18
 Operator : KP/MD
 Sample : 03044-11
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-07-0-2.0

Quant Time: Jun 08 02:33:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	136907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	235235	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	238883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	107419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	91408	50.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.880%	
35) Dibromofluoromethane	7.716	113	81520	54.530	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.060%	
50) Toluene-d8	10.179	98	292595	51.311	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.620%	
62) 4-Bromofluorobenzene	12.483	95	95765	49.261	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 98.520%	
Target Compounds						
16) Acetone	3.954	43	14626	23.588	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
Data File : VY014060.D
Acq On : 07 Jun 2023 15:18
Operator : KP/MD
Sample : 03044-11
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-07-0-2.0

Quant Time: Jun 08 02:33:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

