

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
 Data File : VY014063.D
 Acq On : 07 Jun 2023 16:29
 Operator : KP/MD
 Sample : 03044-15
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-08-5.0-7.0

Quant Time: Jun 08 02:33:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	124823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	209135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	212013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	94348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	84104	50.899	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.800%	
35) Dibromofluoromethane	7.716	113	72866	54.825	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.640%	
50) Toluene-d8	10.179	98	258244	50.939	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.880%	
62) 4-Bromofluorobenzene	12.483	95	81817	47.338	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.680%	
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	8800	15.566	ug/l	93
25) 2-Butanone	6.984	43	4641	6.590	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
Data File : VY014063.D
Acq On : 07 Jun 2023 16:29
Operator : KP/MD
Sample : 03044-15
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-08-5.0-7.0

Quant Time: Jun 08 02:33:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

