

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
 Data File : VY014064.D
 Acq On : 07 Jun 2023 16:52
 Operator : KP/MD
 Sample : 03044-01RE
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02-0-2.0RE

Quant Time: Jun 08 02:34:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	107814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	173344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	162877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	63075	44.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.380%
35) Dibromofluoromethane	7.716	113	57070	51.805	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.620%
50) Toluene-d8	10.179	98	211029	50.221	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.483	95	59871	41.793	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.580%
Target Compounds						
16) Acetone	3.960	43	6382	13.070	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060723\
Data File : VY014064.D
Acq On : 07 Jun 2023 16:52
Operator : KP/MD
Sample : 03044-01RE
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02-0-2.0RE

Quant Time: Jun 08 02:34:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

