

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005015.D
 Acq On : 08 Jun 2021 12:20
 Operator : SY/MD
 Sample : M2612-20
 Misc : 15.19G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(82-86)

Quant Time: Jun 09 03:48:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	3420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	4040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1247	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1726	109.721	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	219.440%#
35) Dibromofluoromethane	7.722	113	721	39.185	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.360%
50) Toluene-d8	10.185	98	4616	60.405	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	120.800%
62) 4-Bromofluorobenzene	12.489	95	1390	52.889	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005015.D
Acq On : 08 Jun 2021 12:20
Operator : SY/MD
Sample : M2612-20
Misc : 15.19G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(82-86)

Quant Time: Jun 09 03:48:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

