

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005017.D
 Acq On : 08 Jun 2021 13:07
 Operator : SY/MD
 Sample : M2626-04RE
 Misc : 6.94G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12-DUP-01RE

Quant Time: Jun 09 03:49:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	156698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	280094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	224464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	59857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	87246	52.667	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.340%	
35) Dibromofluoromethane	7.722	113	84202	55.876	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.760%	
50) Toluene-d8	10.185	98	339017	54.169	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.340%	
62) 4-Bromofluorobenzene	12.483	95	75397	35.029	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 70.060%	
Target Compounds						Qvalue
16) Acetone	3.960	43	7967	19.978	ug/l	95
17) Carbon Disulfide	4.204	76	50526	10.203	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005017.D
Acq On : 08 Jun 2021 13:07
Operator : SY/MD
Sample : M2626-04RE
Misc : 6.94G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12-DUP-01RE

Quant Time: Jun 09 03:49:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

