

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005021.D
 Acq On : 08 Jun 2021 14:40
 Operator : SY/MD
 Sample : M2625-12
 Misc : 6.62G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(182-186)

Quant Time: Jun 09 03:50:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	161628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	298394	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	255453	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	73366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	102070	59.736	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 119.480%		
35) Dibromofluoromethane	7.728	113	95523	59.501	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 119.000%		
50) Toluene-d8	10.185	98	362235	54.329	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 108.660%		
62) 4-Bromofluorobenzene	12.483	95	89774	39.151	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 78.300%		
Target Compounds						
						Qvalue
16) Acetone	3.960	43	10585	25.734	ug/l	96
17) Carbon Disulfide	4.198	76	66331	12.986	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005021.D
Acq On : 08 Jun 2021 14:40
Operator : SY/MD
Sample : M2625-12
Misc : 6.62G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(182-186)

Quant Time: Jun 09 03:50:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

