

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005026.D
 Acq On : 08 Jun 2021 16:36
 Operator : SY/MD
 Sample : M2625-17
 Misc : 6.97G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(202-206)

Quant Time: Jun 09 03:51:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	163508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	294553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	280141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	127454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	100212	57.974	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.940%
35) Dibromofluoromethane	7.728	113	88813	56.043	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.080%
50) Toluene-d8	10.185	98	365611	55.550	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.100%
62) 4-Bromofluorobenzene	12.483	95	123951	54.760	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.520%
Target Compounds						Qvalue
11) Tert butyl alcohol	4.991	59	62768	315.752	ug/l	97
16) Acetone	3.966	43	266621	640.746	ug/l	99
25) 2-Butanone	7.002	43	14909	21.308	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005026.D
Acq On : 08 Jun 2021 16:36
Operator : SY/MD
Sample : M2625-17
Misc : 6.97G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(202-206)

Quant Time: Jun 09 03:51:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

