

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005032.D
 Acq On : 08 Jun 2021 18:55
 Operator : SY/MD
 Sample : M2637-03
 Misc : 7.11G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(238-242)

Quant Time: Jun 09 03:52:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	162712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	290171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	271748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	95719	55.646	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.300%
35) Dibromofluoromethane	7.728	113	86626	55.488	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.980%
50) Toluene-d8	10.185	98	360676	55.628	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.260%
62) 4-Bromofluorobenzene	12.483	95	118318	53.061	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.120%
Target Compounds						
17) Carbon Disulfide	4.198	76	5771	1.122	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005032.D
Acq On : 08 Jun 2021 18:55
Operator : SY/MD
Sample : M2637-03
Misc : 7.11G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(238-242)

Quant Time: Jun 09 03:52:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

