

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005034.D
 Acq On : 08 Jun 2021 19:41
 Operator : SY/MD
 Sample : M2637-05
 Misc : 5.44G/5ML/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 18-B12(246-248)

Quant Time: Jun 09 03:53:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	164075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	307285	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	301994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	136553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	110003	63.418	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.840%	
35) Dibromofluoromethane	7.728	113	97781	59.145	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 118.300%	
50) Toluene-d8	10.185	98	389330	56.703	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 113.400%	
62) 4-Bromofluorobenzene	12.483	95	132799	56.238	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.480%	
Target Compounds						
16) Acetone	3.966	43	4251	10.181	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005034.D
Acq On : 08 Jun 2021 19:41
Operator : SY/MD
Sample : M2637-05
Misc : 5.44G/5ML/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
18-B12(246-248)

Quant Time: Jun 09 03:53:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

