

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005037.D
 Acq On : 08 Jun 2021 20:50
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:54:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.454	0.420	7.5	67	0.00
3 P	Chloromethane	0.620	0.633	-2.1	72	0.00
4 C	Vinyl Chloride	0.596	0.589	1.2#	69	0.00
5 T	Bromomethane	0.338	0.330	2.4	70	0.00
6 T	Chloroethane	0.357	0.366	-2.5	71	0.00
7 T	Trichlorofluoromethane	0.844	0.857	-1.5	71	0.00
8 T	Diethyl Ether	0.328	0.369	-12.5	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.530	-2.9	73	0.00
10 T	Methyl Iodide	0.491	0.559	-13.8	77	0.00
11 T	Tert butyl alcohol	0.061	0.047	23.0	55	0.01
12 CM	1,1-Dichloroethene	0.514	0.526	-2.3#	72	0.00
13 T	Acrolein	0.062	0.055	11.3	63	0.00
14 T	Allyl chloride	0.937	0.996	-6.3	73	0.00
15 T	Acrylonitrile	0.154	0.172	-11.7	73	0.00
16 T	Acetone	0.127	0.131	-3.1	71	0.00
17 T	Carbon Disulfide	1.580	1.571	0.6	68	0.00
18 T	Methyl Acetate	0.408	0.453	-11.0	76	0.00
19 T	Methyl tert-butyl Ether	1.323	1.539	-16.3	78	0.00
20 T	Methylene Chloride	0.741	0.730	1.5	81	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.620	-7.5	74	0.00
22 T	Diisopropyl ether	1.720	2.023	-17.6	80	0.00
23 T	Vinyl Acetate	1.251	1.478	-18.1	77	0.00
24 P	1,1-Dichloroethane	1.035	1.169	-12.9	78	0.00
25 T	2-Butanone	0.214	0.235	-9.8	71	0.00
26 T	2,2-Dichloropropane	0.875	0.854	2.4	69	0.00
27 T	cis-1,2-Dichloroethene	0.632	0.708	-12.0	78	0.00
28 T	Bromochloromethane	0.462	0.524	-13.4	75	0.00
29 T	Tetrahydrofuran	0.136	0.152	-11.8	72	0.00
30 C	Chloroform	0.982	1.140	-16.1#	79	0.00
31 T	Cyclohexane	0.931	0.908	2.5	72	0.00
32 T	1,1,1-Trichloroethane	0.861	0.950	-10.3	75	0.00
33 S	1,2-Dichloroethane-d4	0.529	0.560	-5.9	70	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.269	0.272	-1.1	70	0.00
36 T	1,1-Dichloropropene	0.489	0.491	-0.4	74	0.00
37 T	Ethyl Acetate	0.322	0.340	-5.6	77	0.00
38 T	Carbon Tetrachloride	0.426	0.453	-6.3	76	0.00
39 T	Methylcyclohexane	0.541	0.525	3.0	72	0.00
40 TM	Benzene	1.400	1.511	-7.9	78	0.00
41 T	Methacrylonitrile	0.177	0.194	-9.6	75	0.00
42 TM	1,2-Dichloroethane	0.404	0.440	-8.9	79	0.00
43 T	Isopropyl Acetate	0.578	0.619	-7.1	75	0.00
44 TM	Trichloroethene	0.366	0.389	-6.3	78	0.00
45 C	1,2-Dichloropropane	0.366	0.400	-9.3#	81	0.00
46 T	Dibromomethane	0.196	0.221	-12.8	82	0.00
47 T	Bromodichloromethane	0.459	0.514	-12.0	82	0.00
48 T	Methyl methacrylate	0.263	0.279	-6.1	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
 Data File : VY005037.D
 Acq On : 08 Jun 2021 20:50
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 03:54:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	57	0.00
50 S	Toluene-d8	1.117	1.133	-1.4	71	0.00
51 T	4-Methyl-2-Pentanone	0.303	0.335	-10.6	77	0.00
52 CM	Toluene	0.847	0.925	-9.2#	79	0.00
53 T	t-1,3-Dichloropropene	0.509	0.554	-8.8	78	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.635	-9.9	80	0.00
55 T	1,1,2-Trichloroethane	0.272	0.308	-13.2	82	0.00
56 T	Ethyl methacrylate	0.389	0.448	-15.2	79	0.00
57 T	1,3-Dichloropropane	0.484	0.552	-14.0	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.197	0.215	-9.1	78	0.00
59 T	2-Hexanone	0.205	0.229	-11.7	76	0.00
60 T	Dibromochloromethane	0.309	0.359	-16.2	83	0.00
61 T	1,2-Dibromoethane	0.261	0.296	-13.4	81	0.00
62 S	4-Bromofluorobenzene	0.384	0.386	-0.5	74	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.418	0.429	-2.6	78	0.00
65 PM	Chlorobenzene	0.991	1.043	-5.2	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.353	0.384	-8.8	83	0.00
67 C	Ethyl Benzene	1.831	1.891	-3.3#	79	0.00
68 T	m/p-Xylenes	0.686	0.704	-2.6	78	0.00
69 T	o-Xylene	0.639	0.673	-5.3	80	0.00
70 T	Styrene	1.064	1.156	-8.6	81	0.00
71 P	Bromoform	0.208	0.231	-11.1	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.777	3.775	0.1	78	0.00
74 T	N-amyl acetate	1.219	1.303	-6.9	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.758	0.813	-7.3	83	0.00
76 T	1,2,3-Trichloropropane	0.548	0.573	-4.6	80	0.00
77 T	Bromobenzene	0.844	0.882	-4.5	80	0.00
78 T	n-propylbenzene	4.553	4.565	-0.3	79	0.00
79 T	2-Chlorotoluene	2.538	2.573	-1.4	79	0.00
80 T	1,3,5-Trimethylbenzene	3.116	3.154	-1.2	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.298	0.293	1.7	74	0.00
82 T	4-Chlorotoluene	2.593	2.625	-1.2	79	0.00
83 T	tert-Butylbenzene	2.636	2.684	-1.8	78	0.00
84 T	1,2,4-Trimethylbenzene	3.097	3.156	-1.9	78	0.00
85 T	sec-Butylbenzene	3.837	3.796	1.1	78	0.00
86 T	p-Isopropyltoluene	3.292	3.288	0.1	78	0.00
87 T	1,3-Dichlorobenzene	1.606	1.666	-3.7	81	0.00
88 T	1,4-Dichlorobenzene	1.602	1.641	-2.4	82	0.00
89 T	n-Butylbenzene	3.166	3.098	2.1	77	0.00
90 T	Hexachloroethane	0.489	0.515	-5.3	84	0.00
91 T	1,2-Dichlorobenzene	1.419	1.515	-6.8	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.138	-9.5	84	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.917	-2.9	79	0.00
94 T	Hexachlorobutadiene	0.470	0.462	1.7	74	0.00
95 T	Naphthalene	1.960	2.247	-14.6	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060821\
Data File : VY005037.D
Acq On : 08 Jun 2021 20:50
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 03:54:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.785	0.848	-8.0	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6