

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY061020\
 Data File : VY002876.D
 Acq On : 10 Jun 2020 12:07
 Operator : SY/MD
 Sample : L2810-01
 Misc : 7.10G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-02-060920

Quant Time: Jun 11 02:21:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 13:13:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	239940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	389217	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	357919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	166103	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	120491	57.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.68%	
35) Dibromofluoromethane	7.72	113	119034	52.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.36%	
50) Toluene-d8	10.18	98	456067	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.48	95	145752	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
Target Compounds						
16) Acetone	3.96	43	11062	22.644	ug/l	90
20) Methylene Chloride	4.73	84	17913	7.503	ug/l	90
25) 2-Butanone	7.00	43	3568	4.688	ug/l	95
68) m/p-Xylenes	11.70	106	4899	1.030	ug/l	99
80) 1,3,5-Trimethylbenzene	12.81	105	9733	1.095	ug/l	96
84) 1,2,4-Trimethylbenzene	13.12	105	14347	1.605	ug/l	95
95) Naphthalene	15.23	128	104970	14.588	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY061020\
Data File : VY002876.D
Acq On : 10 Jun 2020 12:07
Operator : SY/MD
Sample : L2810-01
Misc : 7.10G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-02-060920

Quant Time: Jun 11 02:21:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
Quant Title : SW846 8260
QLast Update : Wed May 27 13:13:33 2020
Response via : Initial Calibration

