

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
 Data File : VY005073.D
 Acq On : 10 Jun 2021 12:34
 Operator : SY/MD
 Sample : M2666-07
 Misc : 4.69G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-TO-C-6

Quant Time: Jun 11 05:04:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	51856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	80048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	66049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	26980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	24739	45.127	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.260%
35) Dibromofluoromethane	7.728	113	24861	57.727	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.460%
50) Toluene-d8	10.185	98	97788	54.672	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.340%
62) 4-Bromofluorobenzene	12.483	95	25583	41.589	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.180%
Target Compounds						Qvalue
16) Acetone	3.954	43	4507	34.152	ug/l #	88
25) 2-Butanone	6.997	43	1470	6.624	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
Data File : VY005073.D
Acq On : 10 Jun 2021 12:34
Operator : SY/MD
Sample : M2666-07
Misc : 4.69G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-TO-C-6

Quant Time: Jun 11 05:04:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

