

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
 Data File : VY005082.D
 Acq On : 10 Jun 2021 17:15
 Operator : SY/MD
 Sample : M2672-05
 Misc : 11.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 17-B6(24-28)

Quant Time: Jun 11 05:07:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	190692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	359129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	347905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	157933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	121413	60.226	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.460%
35) Dibromofluoromethane	7.728	113	110832	57.362	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.720%
50) Toluene-d8	10.185	98	440699	54.919	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.840%
62) 4-Bromofluorobenzene	12.483	95	152375	55.213	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
Data File : VY005082.D
Acq On : 10 Jun 2021 17:15
Operator : SY/MD
Sample : M2672-05
Misc : 11.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
17-B6(24-28)

Quant Time: Jun 11 05:07:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

