

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
 Data File : VY005089.D
 Acq On : 10 Jun 2021 19:57
 Operator : SY/MD
 Sample : M2673-02
 Misc : 8.79G/5ML/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 17-B6(92-96)

Quant Time: Jun 11 05:08:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	173017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	314825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	312349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	115733	63.273	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.540%	
35) Dibromofluoromethane	7.728	113	98226	57.992	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.980%	
50) Toluene-d8	10.185	98	393547	55.945	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.880%	
62) 4-Bromofluorobenzene	12.483	95	136597	56.461	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.920%	
Target Compounds						
16) Acetone	3.966	43	4277	9.714	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
Data File : VY005089.D
Acq On : 10 Jun 2021 19:57
Operator : SY/MD
Sample : M2673-02
Misc : 8.79G/5ML/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
17-B6(92-96)

Quant Time: Jun 11 05:08:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

