

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
 Data File : VY005101.D
 Acq On : 11 Jun 2021 02:30
 Operator : SY/MD
 Sample : M2672-16
 Misc : 8.56G/5ML/MSVOA_Y/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 17-B6(68-72)

Quant Time: Jun 11 07:07:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	166838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	305210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	299119	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	142098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	110229	62.496	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.000%	
35) Dibromofluoromethane	7.728	113	94311	57.434	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.860%	
50) Toluene-d8	10.185	98	375657	55.084	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.160%	
62) 4-Bromofluorobenzene	12.483	95	132214	56.371	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.740%	
Target Compounds						
16) Acetone	3.954	43	4188	9.864	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
Data File : VY005101.D
Acq On : 11 Jun 2021 02:30
Operator : SY/MD
Sample : M2672-16
Misc : 8.56G/5ML/MSVOA_Y/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
17-B6(68-72)

Quant Time: Jun 11 07:07:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

