

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
 Data File : VY005109.D
 Acq On : 11 Jun 2021 05:59
 Operator : SY/MD
 Sample : M2678-03
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 0207-COMP-G(DISCRETE-CL-07)

Quant Time: Jun 11 07:09:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	121091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	218053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	215130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	74827	58.452	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.900%
35) Dibromofluoromethane	7.734	113	67981	57.947	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.900%
50) Toluene-d8	10.185	98	268000	55.005	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.020%
62) 4-Bromofluorobenzene	12.483	95	90249	53.859	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.720%
Target Compounds						
16) Acetone	3.966	43	4679	15.183	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061021\
Data File : VY005109.D
Acq On : 11 Jun 2021 05:59
Operator : SY/MD
Sample : M2678-03
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
0207-COMP-G(DISCRETE-CL-07)

Quant Time: Jun 11 07:09:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

