

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
 Data File : VY005117.D
 Acq On : 11 Jun 2021 12:24
 Operator : SY/MD
 Sample : M2675-01
 Misc : 6.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BS-RO-COMP

Quant Time: Jun 12 00:35:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	165009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	296887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	288878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	131125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	106616	61.118	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.240%	
35) Dibromofluoromethane	7.728	113	92041	57.623	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	115.240%	
50) Toluene-d8	10.185	98	364087	54.884	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	109.760%	
62) 4-Bromofluorobenzene	12.483	95	121747	53.364	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	106.720%	
Target Compounds						
16) Acetone	3.966	43	8907	21.211	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
Data File : VY005117.D
Acq On : 11 Jun 2021 12:24
Operator : SY/MD
Sample : M2675-01
Misc : 6.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BS-RO-COMP

Quant Time: Jun 12 00:35:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

