

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
 Data File : VY005119.D
 Acq On : 11 Jun 2021 13:11
 Operator : SY/MD
 Sample : M2672-08
 Misc : 9.96G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 17-B6(36-40)

Quant Time: Jun 12 00:35:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	157584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	282070	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	277473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	132477	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	101504	60.929	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.860%
35) Dibromofluoromethane	7.728	113	87611	57.731	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.460%
50) Toluene-d8	10.185	98	349202	55.405	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.820%
62) 4-Bromofluorobenzene	12.483	95	119864	55.298	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
Data File : VY005119.D
Acq On : 11 Jun 2021 13:11
Operator : SY/MD
Sample : M2672-08
Misc : 9.96G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
17-B6(36-40)

Quant Time: Jun 12 00:35:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

