

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
 Data File : VY005113.D
 Acq On : 11 Jun 2021 10:20
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 00:34:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.454	0.490	-7.9	81	0.00
3 P	Chloromethane	0.620	0.626	-1.0	75	0.00
4 C	Vinyl Chloride	0.596	0.611	-2.5#	74	0.00
5 T	Bromomethane	0.338	0.318	5.9	71	0.00
6 T	Chloroethane	0.357	0.341	4.5	69	0.00
7 T	Trichlorofluoromethane	0.844	0.968	-14.7	84	0.00
8 T	Diethyl Ether	0.328	0.340	-3.7	75	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.558	-8.3	80	0.00
10 T	Methyl Iodide	0.491	0.590	-20.2	84	0.00
11 T	Tert butyl alcohol	0.061	0.054	11.5	66	0.00
12 CM	1,1-Dichloroethene	0.514	0.554	-7.8#	79	0.00
13 T	Acrolein	0.062	0.056	9.7	67	0.00
14 T	Allyl chloride	0.937	0.887	5.3	68	0.00
15 T	Acrylonitrile	0.154	0.158	-2.6	70	0.00
16 T	Acetone	0.127	0.112	11.8	64	0.00
17 T	Carbon Disulfide	1.580	1.607	-1.7	73	0.00
18 T	Methyl Acetate	0.408	0.402	1.5	70	0.00
19 T	Methyl tert-butyl Ether	1.323	1.403	-6.0	74	0.00
20 T	Methylene Chloride	0.741	0.675	8.9	78	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.602	-4.3	75	0.00
22 T	Diisopropyl ether	1.720	1.641	4.6	68	0.00
23 T	Vinyl Acetate	1.251	1.199	4.2	66	0.00
24 P	1,1-Dichloroethane	1.035	1.050	-1.4	73	0.00
25 T	2-Butanone	0.214	0.197	7.9	63	0.00
26 T	2,2-Dichloropropane	0.875	0.964	-10.2	81	0.00
27 T	cis-1,2-Dichloroethene	0.632	0.674	-6.6	78	0.00
28 T	Bromochloromethane	0.462	0.406	12.1	61	0.00
29 T	Tetrahydrofuran	0.136	0.128	5.9	64	0.00
30 C	Chloroform	0.982	1.069	-8.9#	78	0.00
31 T	Cyclohexane	0.931	0.854	8.3	71	0.00
32 T	1,1,1-Trichloroethane	0.861	0.977	-13.5	81	0.00
33 S	1,2-Dichloroethane-d4	0.529	0.510	3.6	67	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.269	0.273	-1.5	66	0.00
36 T	1,1-Dichloropropene	0.489	0.533	-9.0	75	0.00
37 T	Ethyl Acetate	0.322	0.308	4.3	66	0.00
38 T	Carbon Tetrachloride	0.426	0.535	-25.6#	85	0.00
39 T	Methylcyclohexane	0.541	0.588	-8.7	76	0.00
40 TM	Benzene	1.400	1.506	-7.6	74	0.00
41 T	Methacrylonitrile	0.177	0.156	11.9	57	0.00
42 TM	1,2-Dichloroethane	0.404	0.448	-10.9	76	0.00
43 T	Isopropyl Acetate	0.578	0.558	3.5	64	0.00
44 TM	Trichloroethene	0.366	0.425	-16.1	81	0.00
45 C	1,2-Dichloropropane	0.366	0.365	0.3#	70	0.00
46 T	Dibromomethane	0.196	0.212	-8.2	75	0.00
47 T	Bromodichloromethane	0.459	0.514	-12.0	78	0.00
48 T	Methyl methacrylate	0.263	0.251	4.6	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
 Data File : VY005113.D
 Acq On : 11 Jun 2021 10:20
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 00:34:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	71	0.00
50 S	Toluene-d8	1.117	1.122	-0.4	67	0.00
51 T	4-Methyl-2-Pentanone	0.303	0.286	5.6	62	0.00
52 CM	Toluene	0.847	0.941	-11.1#	76	0.00
53 T	t-1,3-Dichloropropene	0.509	0.549	-7.9	73	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.626	-8.3	74	0.00
55 T	1,1,2-Trichloroethane	0.272	0.301	-10.7	76	0.00
56 T	Ethyl methacrylate	0.389	0.423	-8.7	70	0.00
57 T	1,3-Dichloropropane	0.484	0.515	-6.4	73	0.00
58 T	2-Chloroethyl Vinyl ether	0.197	0.182	7.6	62	0.00
59 T	2-Hexanone	0.205	0.199	2.9	62	0.00
60 T	Dibromochloromethane	0.309	0.360	-16.5	78	0.00
61 T	1,2-Dibromoethane	0.261	0.288	-10.3	75	0.00
62 S	4-Bromofluorobenzene	0.384	0.377	1.8	68	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	71	0.00
64 T	Tetrachloroethene	0.418	0.483	-15.6	80	0.00
65 PM	Chlorobenzene	0.991	1.094	-10.4	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.353	0.404	-14.4	80	0.00
67 C	Ethyl Benzene	1.831	2.017	-10.2#	77	0.00
68 T	m/p-Xylenes	0.686	0.763	-11.2	77	0.00
69 T	o-Xylene	0.639	0.723	-13.1	78	0.00
70 T	Styrene	1.064	1.217	-14.4	78	0.00
71 P	Bromoform	0.208	0.246	-18.3	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
73 T	Isopropylbenzene	3.777	3.992	-5.7	79	0.00
74 T	N-amyl acetate	1.219	1.062	12.9	61	0.00
75 P	1,1,2,2-Tetrachloroethane	0.758	0.732	3.4	71	0.00
76 T	1,2,3-Trichloropropane	0.548	0.572	-4.4	76	0.00
77 T	Bromobenzene	0.844	0.913	-8.2	79	0.00
78 T	n-propylbenzene	4.553	4.793	-5.3	79	0.00
79 T	2-Chlorotoluene	2.538	2.605	-2.6	76	0.00
80 T	1,3,5-Trimethylbenzene	3.116	3.388	-8.7	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.298	0.292	2.0	71	0.00
82 T	4-Chlorotoluene	2.593	2.694	-3.9	78	0.00
83 T	tert-Butylbenzene	2.636	2.932	-11.2	82	0.00
84 T	1,2,4-Trimethylbenzene	3.097	3.357	-8.4	80	0.00
85 T	sec-Butylbenzene	3.837	4.160	-8.4	82	0.00
86 T	p-Isopropyltoluene	3.292	3.678	-11.7	83	0.00
87 T	1,3-Dichlorobenzene	1.606	1.750	-9.0	82	0.00
88 T	1,4-Dichlorobenzene	1.602	1.761	-9.9	84	0.00
89 T	n-Butylbenzene	3.166	3.459	-9.3	82	0.00
90 T	Hexachloroethane	0.489	0.543	-11.0	85	0.00
91 T	1,2-Dichlorobenzene	1.419	1.580	-11.3	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.135	-7.1	78	0.00
93 T	1,2,4-Trichlorobenzene	0.891	1.071	-20.2	88	0.00
94 T	Hexachlorobutadiene	0.470	0.595	-26.6#	91	0.00
95 T	Naphthalene	1.960	2.374	-21.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061121\
Data File : VY005113.D
Acq On : 11 Jun 2021 10:20
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 00:34:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.785	0.966	-23.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6