



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061224\
Data File : VY018566.D
Acq On : 12 Jun 2024 10:41
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Title : SW846 8260
Last Update : Tue Jun 04 07:11:56 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	26.1	18473	PASS
75	95	30	60	59.1	41931	PASS
95	95	100	100	100.0	70909	PASS
96	95	5	9	6.1	4357	PASS
173	174	0.00	2	0.8	420	PASS
174	95	50	100	75.4	53485	PASS
175	174	5	9	7.1	3803	PASS
176	174	95	101	96.8	51747	PASS
177	176	5	9	6.3	3248	PASS