

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061423\
 Data File : VY014163.D
 Acq On : 14 Jun 2023 19:00
 Operator : KP/MD
 Sample : 03223-05
 Misc : 4.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-SB-04-GRAB-6

Quant Time: Jun 15 01:45:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	145256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	251839	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	246633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90356	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	105967	52.389	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.780%
35) Dibromofluoromethane	7.716	113	87451	48.965	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.940%
50) Toluene-d8	10.179	98	310227	46.462	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.920%
62) 4-Bromofluorobenzene	12.483	95	93336	39.351	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.700%
Target Compounds						
29) Tetrahydrofuran	7.356	42	3420	7.186	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061423\
Data File : VY014163.D
Acq On : 14 Jun 2023 19:00
Operator : KP/MD
Sample : 03223-05
Misc : 4.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-SB-04-GRAB-6

Quant Time: Jun 15 01:45:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

