

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061424\
 Data File : VY018596.D
 Acq On : 14 Jun 2024 15:45
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 14 23:24:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	90697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	174844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	158330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	59314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	52110	57.068	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.140%	
35) Dibromofluoromethane	7.710	113	54171	50.789	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.580%	
50) Toluene-d8	10.173	98	206341	50.306	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.620%	
62) 4-Bromofluorobenzene	12.477	95	57400	44.318	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 88.640%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	19478	9.811	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061424\
Data File : VY018596.D
Acq On : 14 Jun 2024 15:45
Operator : SY/MD
Sample : VIBLK
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jun 14 23:24:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

