

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
 Data File : VY014184.D
 Acq On : 15 Jun 2023 13:09
 Operator : KP/MD
 Sample : 03222-05
 Misc : 10.77g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-GRAB-4-SB-33

Quant Time: Jun 16 02:07:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	164396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	291502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	266316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	119945	52.395	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.800%
35) Dibromofluoromethane	7.722	113	99962	48.355	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.700%
50) Toluene-d8	10.185	98	355688	46.022	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.040%
62) 4-Bromofluorobenzene	12.483	95	91808	33.440	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	66.880%
Target Compounds						
29) Tetrahydrofuran	7.350	42	3080	5.718	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
Data File : VY014184.D
Acq On : 15 Jun 2023 13:09
Operator : KP/MD
Sample : 03222-05
Misc : 10.77g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-GRAB-4-SB-33

Quant Time: Jun 16 02:07:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

