

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
 Data File : VY014186.D
 Acq On : 15 Jun 2023 13:56
 Operator : KP/MD
 Sample : 03222-11
 Misc : 4.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-GRAB-4-SB-26

Quant Time: Jun 16 02:07:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	172393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	304437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	297530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	116723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	125792	52.400	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.800%
35) Dibromofluoromethane	7.716	113	103784	48.070	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.140%
50) Toluene-d8	10.179	98	385787	47.795	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.600%
62) 4-Bromofluorobenzene	12.483	95	117400	40.945	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.880%
Target Compounds						Qvalue
29) Tetrahydrofuran	7.356	42	3616	6.402	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
Data File : VY014186.D
Acq On : 15 Jun 2023 13:56
Operator : KP/MD
Sample : 03222-11
Misc : 4.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-GRAB-4-SB-26

Quant Time: Jun 16 02:07:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

