

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
 Data File : VY014191.D
 Acq On : 15 Jun 2023 15:53
 Operator : KP/MD
 Sample : 03223-11
 Misc : 8.35g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-SB-05-GRAB-6

Quant Time: Jun 16 02:09:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	172938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	305571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	274321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	82215	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	133247	55.331	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.660%
35) Dibromofluoromethane	7.716	113	106378	49.089	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.180%
50) Toluene-d8	10.179	98	372108	45.930	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.860%
62) 4-Bromofluorobenzene	12.483	95	93732	32.569	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.140%
Target Compounds						
29) Tetrahydrofuran	7.344	42	2854	5.037	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
Data File : VY014191.D
Acq On : 15 Jun 2023 15:53
Operator : KP/MD
Sample : 03223-11
Misc : 8.35g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-SB-05-GRAB-6

Quant Time: Jun 16 02:09:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

