

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
 Data File : VY014196.D
 Acq On : 15 Jun 2023 17:50
 Operator : KP/MD
 Sample : 03235-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A-MANHOLE

Quant Time: Jun 16 02:11:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	35611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	56622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	56959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	25114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	22732	45.841	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.680%
35) Dibromofluoromethane	7.716	113	20032	49.887	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.780%
50) Toluene-d8	10.179	98	68859	45.868	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.740%
62) 4-Bromofluorobenzene	12.483	95	21867	41.005	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.000%
Target Compounds						
16) Acetone	3.942	43	1849	12.370	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
Data File : VY014196.D
Acq On : 15 Jun 2023 17:50
Operator : KP/MD
Sample : 03235-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A-MANHOLE

Quant Time: Jun 16 02:11:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

