

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
 Data File : VY014199.D
 Acq On : 15 Jun 2023 19:00
 Operator : KP/MD
 Sample : 03139-02
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jun 16 02:12:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	227284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	409940	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	440542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	211650	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	164555	51.993	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.980%
35) Dibromofluoromethane	7.716	113	138024	47.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.960%
50) Toluene-d8	10.179	98	525687	48.366	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.740%
62) 4-Bromofluorobenzene	12.483	95	192505	49.860	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061523\
Data File : VY014199.D
Acq On : 15 Jun 2023 19:00
Operator : KP/MD
Sample : 03139-02
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jun 16 02:12:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

