

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005150.D
 Acq On : 17 Jun 2021 12:27
 Operator : SY/MD
 Sample : VY0617SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0617SBL01

Quant Time: Jun 18 02:59:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	166811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	298835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	284419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	130962	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	99372	61.611	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.220%
35) Dibromofluoromethane	7.728	113	90229	70.246	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	140.500%
50) Toluene-d8	10.185	98	360604	56.641	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.280%
62) 4-Bromofluorobenzene	12.483	95	120079	56.081	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005150.D
Acq On : 17 Jun 2021 12:27
Operator : SY/MD
Sample : VY0617SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0617SBL01

Quant Time: Jun 18 02:59:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

