

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005155.D
 Acq On : 17 Jun 2021 14:37
 Operator : SY/MD
 Sample : M2749-02
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PX-02 14.5-15.0

Quant Time: Jun 18 03:01:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	142252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	251102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	246741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	116705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	90163	65.553	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.100%
35) Dibromofluoromethane	7.728	113	78346	72.816	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	145.640%
50) Toluene-d8	10.184	98	306987	57.385	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.780%
62) 4-Bromofluorobenzene	12.483	95	104686	58.186	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.380%
Target Compounds						
16) Acetone	3.966	43	4337	15.503	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005155.D
Acq On : 17 Jun 2021 14:37
Operator : SY/MD
Sample : M2749-02
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PX-02 14.5-15.0

Quant Time: Jun 18 03:01:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

