

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005158.D
 Acq On : 17 Jun 2021 15:47
 Operator : SY/MD
 Sample : M2749-05
 Misc : 6.24G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PP-01 4.5-5.0

Quant Time: Jun 18 03:02:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	153980	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	279733	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	276206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	130218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	100052	67.202	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.400%
35) Dibromofluoromethane	7.728	113	85970	71.622	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	143.240%
50) Toluene-d8	10.185	98	343013	57.557	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.120%
62) 4-Bromofluorobenzene	12.483	95	118843	59.294	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.580%
Target Compounds						
16) Acetone	3.960	43	4141	13.675	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005158.D
Acq On : 17 Jun 2021 15:47
Operator : SY/MD
Sample : M2749-05
Misc : 6.24G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PP-01_4.5-5.0

Quant Time: Jun 18 03:02:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

