

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005163.D
 Acq On : 17 Jun 2021 17:43
 Operator : SY/MD
 Sample : M2748-07
 Misc : 5.91G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8A

Quant Time: Jun 18 03:03:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	148662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	271266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	268711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	132915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	100915	70.206	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	140.420%	
35) Dibromofluoromethane	7.728	113	84961	73.120	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	146.240%	
50) Toluene-d8	10.185	98	335873	58.118	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	116.240%	
62) 4-Bromofluorobenzene	12.483	95	118373	60.903	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	121.800%	
Target Compounds						
20) Methylene Chloride	4.728	84	48678	27.252	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005163.D
Acq On : 17 Jun 2021 17:43
Operator : SY/MD
Sample : M2748-07
Misc : 5.91G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8A

Quant Time: Jun 18 03:03:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

